



पत्र संख्या: ०८१/०८२
चलानी नम्बर:

शिवालय गाउँपालिका
Shibalaya Rural Municipality
गाउँ कार्यपालिकाको कार्यालय
Office of The Rural Municipal Executive
सिमा, जाजरकोट
Sima, Jajarkot




कर्णाली प्रदेश, नेपाल
Karnali Province, Nepal

दर रेट पेश गर्ने सम्बन्धमा

(प्रथम पटक सूचना प्रकाशित मिति २०८१/१२/२१)

प्रस्तुत बिषयमा शिवालय गाउँपालिकाको चालू आ.व. २०८१/०८२ को स्वीकृत बजेट बाट यस पालिका भित्र संचालित वर्थिङ सेन्टरहरु ब्यवस्थापनका लागि आवश्यक पर्ने मेडिकल उपकरणहरु खरिद गर्नुपर्ने भएकोले इच्छुक फर्म वा संस्थाहरुले सूचना भएको मितिले ७ दिन भित्र यसै सूचनाका साथ संलग्न BOQ भरी आफ्नो शिलवन्दी दरभाउ पत्र पेश गर्न यो सूचना प्रकाशन गरिन्छ ।


गोपाल शाही

प्रमुख प्रशासकिय अधिकृत
प्रमुख प्रशासकिय अधिकृत

तपशिल

१. निवेदन

२. भ्याट प्यान दर्ता प्रमाण प्रतिलिपि

३. ब्यवसाय दर्ता प्रमाण प्रतिलिपि

४. आ.व. २०८०/०८१ को कर चुक्ता प्रमाण पत्र प्रतिलिपि

५. संलग्न BOQ फारम



शिवालय गाउँपालिका
गाउँ कार्यपालिकाको कार्यालय

चिमा, जलजकोट
कर्णाली प्रदेश, नेपाल
BOQ



S.N.	Criteria	Descriptions	इकाई	परिमाण	प्रति इकाई दर	जम्मा रकम
1	Scope of Work	Procurement of medical equipment.				
1.1	Baby Warmer	A microprocessor - controlled temperature, open care radiant warmer system able to maintain appropriate body temperature for premature babies.				
i.	Control Panel	Automatic, manual, baby modes (Servo Mode), Changeable from °C & °F or vice versa by push button, a full panel of alarm with both audible and visual indicators. Set temperature range: 30°C - 38°C (adjustable in increments of 0.1°C)				
ii.	Sensors	Air thermal heat sensor, alarms, skin temperature sensor				
iii.	Heater	Silica / Quartz / Ceramic Heater, up to 650W, heater can be swiveled in either direction.				
iv.	Baby Bed / Pad	Spacious bed with drop down / fold downside panels for maximum access to the patient, hypoallergenic foam mattress with waterproof and washable cover, adjustable height, ergonomic design for infant comfort	SET	1		
v.	Accessories, spares and consumables	It shall come with extra: <ul style="list-style-type: none">Air sensor 1 pc.Skin sensor 1 pc.Heating element 1 pc. All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials to be included in the offer.				
vi.	Power supply	AC 220 V ± 10%, 50 Hz, at least 3 meters power cable with appropriate plug				
vii.	Additional Features	<ul style="list-style-type: none">The facility to put IV stands on both side of side bar.Built-in rechargeable battery back-up with autonomy of 2 to 3 hrs of operation except heating facility.				
1.2	Peri Light	Portable LED-based perimeter lighting or task lighting suitable for medical use				
i.	Lamp	Light weight design with easy positioning up to 40° angle, adjustable brightness, min. 50W light output with extra focus				
ii.	Materials	Stainless Steel housing				
iii.	CRI (Color Rendering Index)	≥ 90 for accurate color representation	SET	3		
iv.	Power supply	220 V ± 10%, 50 Hz, at least 3 meters power cable with appropriate plug				
v.	Safety Features	Overheat protection, glare reduction and flicker-free operation				
vi.	Accessories	<ul style="list-style-type: none">Spare lamp: 1 set All standard accessories, consumables and parts required to operate the equipment.				
1.3	Autoclave Drum	Cylindrical containers for sterilizing dressing materials, surgical instruments and other medical items designed to withstand high-pressure & temperature, allow steam to circulate freely through air vents during autoclaving.				
i.	Materials	Austenitic Stainless steel (Composition: approx. 8% to 10% nickel, 18% to 20% chromium) with smooth surface finish.	PCS	4		

प्रमुख प्रशासकीय अधिकारी



शिवालय गाउँपालिका
गाउँ कार्यपालिकाको कार्यालय

सिन्धु, जाजरकोट
कर्णाली प्रदेश, नेपाल
BOQ



ii.	Dimensions	Diameter: Approx. 300 mm Height: Approx. 180 mm to 200 mm Thickness: Approx. 0.6 to 0.7 mm				
iii.	Lid Type	Perforated or solid lid with locking mechanism for secure sealing				
iv.	Handles	<ul style="list-style-type: none">Ergonomic handles for easy transport and handling				
v.	Perforation	Perforated body for effective steam penetration and drainage.				
1.4	Suction Tube	Soft and kink-resistant transparent tube designed to withstand high negative pressure without collapsing and can be used for drainage and suction of fluids, blood, or secretions during procedures.				
i.	Materials	Medical grade PVC or silicone (latex-free)	SET	20		
ii.	Dimensions	Standard Diameter 7 mm (inner) × 12 mm (outer), Length min. 180 cm				
iii.	Connector Type	Universal female connector (Straight or flared)				
iv.	Sterilization	Ethylene Oxide sterilized, single use				
1.5	Emergency Medicine Tray	Light weight rectangular tray with multiple sections for organizing medications, syringes and small instruments.				
i.	Materials	Stainless steel with smooth surface finish.	PCS	4		
ii.	Dimensions	Standard Size: 12" × 15" × 2"				
iii.	Lid Type	Operational lid with central handle for secure closure				
iv.	Sterilization	Compatible with autoclave or chemical sterilization methods				
1.6	Doppler	Compact and portable equipment ensuring safety, accuracy and ease of use for monitoring fetal heart rates				
i.	Probe	Convex probe with the nominal working frequency of 2 MHz (standard for fetal heart rate detection)				
ii.	Imaging mode	3D / 4D with less than 10 mW/cm ² ultrasound intensity safe for foetal use.				
iii.	Heart Rate Range	50 – 240 bpm with ± 1 bpm accuracy	SET	2		
iv.	Display	Digital LED screen				
v.	Power supply	Rechargeable battery with min. 8 hours of continuous operation after full charge				
vi.	Additional Features	<ul style="list-style-type: none">Built-in speaker, earphone compatibility, gel compatibility, and carrying case.				
1.7	Postnatal Care Bed	A durable frame with comfortable and adjustable position bed allowing easy mobility using 360° swivel castors with individual brakes				
i.	Dimensions	<ul style="list-style-type: none">Size (buffer to buffer) (L×W×H): 1800 – 2000 mm × 820 – 900 mm × 600 – 750 mmPlatform (L×W): 1700 – 1900 mm × 720 – 820 mmMattress platform (L×W): 1600 – 1800 mm × 700 – 800 mm	SET	1		
ii.	Materials	Epoxy powder-coated steel frame with ABS head and footboards				
iii.	Adjustable Positions	Back rest: 0° to 65° Leg rest: 0° to 45° Trendelenburg: 10° to 12° Reverse Trendelenburg: 8° Maneuvered by the screw mechanism.				

प्रमुख प्रशासकीय अधिकारी



शिवालय गाउँपालिका
गाउँ कार्यपालिकाको कार्यालय

सिमा, जाजरकोट
कर्णाली प्रदेश, नेपाल
BOQ



iii.	Side Railing	Collapsible, tubular & telescopic stainless steel side railing for patient safety				
v.	Mattress	Two section mattress with high density foam of 40 kg/m ³ with PVC, waterproof, washable cover				
vi.	Pillow	Standard size with Rixin cover				
vii.	Accessories	Telescopic IV Rod, S.S. saline rod, urine bag holder, and utility hooks, 4 rod receptacles at the 4 corners. All standard accessories & parts required for the proper operation of the above item shall be included in the offer				
2	Standards & Safety Requirements	<ul style="list-style-type: none">• Must submit ISO 9001 or ISO13485 for Medical Devices• CE or USFDA approved product certificate.• Shall meet IEC -60601-1-2 general requirement of safety of electrically operated equipment.				
3	Operating environment	The equipment offered shall be designed to store and to operate normally under the conditions of the purchaser's country. The conditions include climate, temperature, humidity power supply etc.				
4	Warranty & Service	<ul style="list-style-type: none">• Comprehensive warranty for 1 year after acceptance.• During the warranty period, the supplier must ensure corrective / breakdown maintenance whenever required.				
5	Instructions & Manuals	One copy of user manuals written in English must be provided				
6	User Training	Must provide user training (including how to use and maintain the equipment)				
7	Delivery, Installation & Commissioning	Supplier must deliver, install and commission the equipment on the designated site as per Shivalaya Rural Municipality Office, Sima, Jajarkot.				

संस्थाको नाम:

पान नः

संस्थाको छापः

प्रतिनिधिको हस्ताक्षरः

प्रमुख प्रशासकीय अधिकारी